

## EU DECLARATION OF CONFORMITY

according to Medical Device regulation (EU) 2017/745 and according to Low Voltage direktive 2014/35/EU and according to EMC directive 2014/30/EU and according to RoHS directive 2011/65/EU and according to REACH directive 1907/2006/EU and according to Eco-Design framework directive 2009/125/EC and according to WEEE directive 2012/19/EU

We as manufacturer: all flow-med GmbH, Weinbergweg 10, 73773 Aichwald, Germany

SRN: DE-MF-000006336

declare under our reponsibility that the following listed medical devices meet all applicable basic safety and performance requirements of the regulations mentioned above.

Product name: dry-cap uv 4®

912 00000 Item number:

Basic UDI-DI: 4260277010096V

EMDN: Y214580

Intended use: Electronic drying and hygiene station for hearing aids

Product classification: Class I, according to annex VIII, rule 1 and 13 of regulation (EU) 2017/745 (MDR)

Applied specifications/ norms: EN 55022:2010, EN 55024:2010, EN 63000:2018



Valid until: 31.12.2026

Aichwald, 01.01.2025

Place, date

 $C \in$ 



Name Managing Director:

Meiko Panitzsch Flow

medical appliances

med

Weinbergweg 10 73773 Aichwald Germany Signature

> phone +49 711 50463500 fax +49 711 50463501